



**REGIONAL AGRICULTURAL RESEARCH STATION**

**Kerala Agricultural University**

MelePattambi (Post), Palakkad (Dt.), 679 306, Kerala  
Email: [adrptb@kau.in](mailto:adrptb@kau.in), [rarsptb@kau.in](mailto:rarsptb@kau.in) Ph : 0466-2212228

No. D3-1552/2024

Dated: 16.03.2024

QUOTATION NOTICE

Sealed quotations on competition basis are invited for the supply of Rotek Hot Air Oven as per terms and conditions detailed below.

Specification of Rotek Hot Air Oven.

- Rotek Hot Air Oven – Bottom Heated, Thermostatic
- Outer Body Made M S Sheet, Powder Coated
- Inner Chamber Stainless Steel SS 304,
- Temperature Range Ambient +5°C +ve 200°C Accuracy  $\pm 1^\circ\text{C}$
- Size 45 X 45 X 60 cm,
- No of Shelves - 3 Nos
- Watts – 1200, Capacity 125 L
- Digital Temperature controller cum indicator
- Air Circulating Fan

Terms and conditions:-

- 1 The quotation should be in the sealed cover, addressed to the Associate Director of Research, Regional Agricultural Research Station, and Pattambi.
- 2 The envelope should be super scribed as: "Quotation for supply of Hot Air Oven " as per quotation notice no. D3-1552/2024 dtd 16.03.2024.
- 3 The last date for submission of quotation is on 25.03.2024 at 11.00 am. Quotation will be opened at 11.30 am on the same day in the presence of those bidders who are present at that time.
- 4 The rate quoted should be inclusive of all taxes and transportation charges up to this station, if any.
- 5 The lowest rate offered as per the specifications will be accepted if reasonable.
- 6 The item should be supplied within 07 days on receipt of supply order issued from this office.
- 7 The payment will be made through invoice/credit bill system and by NEFT only to the bank account of the supplier.
- 8 The decision of the undersigned will be final in connection with this quotation and all the other terms and conditions of government quotations in force will be applicable in this quotation also.

Sd/-

Associate Director of Research (i/c)

To: Notice Board-Office, Farm, Sales counter, KAU Website  
Cc to: Dr. Thulasi V, Associate Professor.

//Approved for issue//

  
Section Officer